
















## REGULATED SYMBOLS

SYMBOL	SYMBOL TITLE	DESCRIPTION	STANDARD & REFERENCE	STANDARD TITLE
<b>Manufacture</b>				
	Manufacturer	Indicates the medical device manufacturer	ISO 15223-1:2021 5.1.1	Medical devices - Symbols to be used with information to be supplied by the manufacturer
	Date of manufacture	Indicates the date when the medical device was manufactured	ISO 15223-1:2021 5.1.3	Medical devices - Symbols to be used with information to be supplied by the manufacturer
	Use-by date	Indicates the date after which the medical device is not to be used	ISO 15223-1:2021 5.1.4	Medical devices - Symbols to be used with information to be supplied by the manufacturer
	Batch code	Indicates the manufacturer's batch code so that the batch or lot can be identified	ISO 15223-1:2021 5.1.5	Medical devices - Symbols to be used with information to be supplied by the manufacturer
	Catalogue number	Indicates the manufacturer's catalogue number so that the medical device can be identified	ISO 15223-1:2021 5.1.6	Medical devices - Symbols to be used with information to be supplied by the manufacturer
	Country of manufacture	To identify the country of manufacture of products	ISO 15223-1:2021 5.1.11	Medical devices - Symbols to be used with information to be supplied by the manufacturer
<b>Sterility</b>				
	Sterilized using ethylene oxide	Indicates a medical device that has been sterilized using ethylene oxide	ISO 15223-1:2021 5.2.3	Medical devices - Symbols to be used with information to be supplied by the manufacturer
	Do not re-sterilize	Indicates a medical device that is not to be re-sterilized	ISO 15223-1:2021 5.2.6	Medical devices - Symbols to be used with information to be supplied by the manufacturer
	Non-sterile	Indicates a medical device that has not been subjected to a sterilization process	ISO 15223-1:2021 5.2.7	Medical devices - Symbols to be used with information to be supplied by the manufacturer

## REGULATED SYMBOLS

SYMBOL	SYMBOL TITLE	DESCRIPTION	STANDARD & REFERENCE	STANDARD TITLE
<b>Sterility</b>				
	Do not use if package is damaged and consult instructions for use	Indicates that a medical device that should not be used if the package has been damaged or opened and that the user should consult the instructions for use for additional information	ISO 15223-1:2021 5.2.8	Medical devices - Symbols to be used with information to be supplied by the manufacturer
	Single sterile barrier system	Indicates a single sterile barrier system	ISO 15223-1:2021 5.2.11	Medical devices - Symbols to be used with information to be supplied by the manufacturer
<b>Safe Use</b>				
	Do not re-use	Indicates a medical device that is intended for one single use only	ISO 15223-1:2021 5.4.2	Medical devices - Symbols to be used with information to be supplied by the manufacturer
	Consult instructions for use or consult electronic instructions for use	Indicates the need for the user to consult the instructions for use	ISO 15223-1:2021 5.4.3	Medical devices - Symbols to be used with information to be supplied by the manufacturer
	Caution	Indicates that caution is necessary when operating the device or control close to where the symbol is placed, or that the current situation needs operator awareness or operator action in order to avoid undesirable consequences	ISO 15223-1:2021 5.4.4	Medical devices - Symbols to be used with information to be supplied by the manufacturer
	Single patient multiple use	Indicates a medical device that may be used multiple times (multiple procedures) on a single patient	ISO 15223-1:2021 5.4.12	Medical devices - Symbols to be used with information to be supplied by the manufacturer
<b>RxOnly</b>	Rx Only or Prescription only	Indicates a medical device that requires a prescription in the United States	FDA-2013-N-0125-0016	Use of Symbols in Labeling